



Canadian Cardiovascular Society
Leadership. Knowledge. Community.



Canadian Cardiovascular Society
Guidelines on the Use of Cardiac Resynchronization
Therapy: Evidence, Patient Selection and
Implementation

Please visit www.ccs.ca for more information or additional resources.



About This Pocket Guide

This pocket guide is a quick-reference tool that features essential diagnostic and treatment recommendations based on the 2013 CCS Cardiac Resynchronization Therapy (CRT) Guidelines.

These recommendations are intended to provide a reasonable and practical approach to care for specialists and allied health professionals with the duty of bestowing optimal care to patients and families. They are subject to change as scientific knowledge and technology advance and practice patterns evolve. The guidelines are not intended to be a substitute for physicians using their judgement in managing clinical care in consultation with the patient, with appropriate regard to the individual circumstances of the patient, diagnostic and treatment options and available resources. Adherence to these recommendations will not necessarily produce successful outcomes in every case.

Please visit www.ccs.ca for more information or additional resources.



Canadian Cardiovascular Society Guidelines on the Use of Cardiac Resynchronization Therapy: Evidence and Patient Selection

Co-chairs and Authors

Derek V. Exner and Ratika Parkash

Authors

David H. Birnie, Gordon Moe, Bernard Thibault, François Philippon, Jeffrey S. Healey, Anthony S.L. Tang and Eric Larose

Canadian Cardiovascular Society Guidelines on the Use of Cardiac Resynchronization Therapy: Implementation

Co-chairs and Authors

Derek V. Exner and Ratika Parkash


Authors

François Philippon, Miriam Shanks, Bernard Thibault, Jafna Cox, Aaron Low, Vidal Essebag, Jamil Bashir, Gordon Moe, David H. Birnie, Eric Larose, Raymond Yee, Elizabeth Swiggum, Padma Kaul, Damian Redfearn and Anthony S. Tang

Guiding Principles.....	5-10
When to Consider CRT.....	11-12
Role of Imaging.....	13
Atrial Fibrillation.....	14
CRT-P vs. CRT-D.....	15
Lead Placement.....	16
Preimplant Assessment.....	17-18
Procedural Preparation.....	19
Operative Issues.....	20
Device Follow-up.....	21
Health Economics and Accessibility of CRT.....	22



Guiding Principles – Patient Classification



See next page for continuation (CRT Considerations)

Practical tip:

- Reasons for nonuse or prescription of lower doses of heart failure medication should be recorded
- The following should be recorded in the patient's medical record:
 - At a minimum, assessment NYHA functional class; 6-minute hall walk distance, disease-specific HRQoL and cardiopulmonary testing should also be considered
 - QRS duration should be measured from a standard 12-lead ECG and LVEF quantified using a validated assessment method



Guiding Principles – CRT Considerations

Implement adequate medical therapy before CRT

- Guideline-specified doses of ACEi/ARBs, β blockers and mineralocorticoid receptor antagonists

LVEF $\leq 35\%$

LVEF $\leq 45\%$

LVEF $> 35\%$

QRS considerations

• QRS duration < 130 ms

• QRS duration ≥ 130 ms due to LBBB
• In sinus rhythm*

• QRS duration ≥ 150 ms not because of LBBB
• In sinus rhythm*

• Chronic RV pacing or patients who are likely to be chronically paced*
• Signs and/or symptoms of HF

Insufficient evidence for CRT

CRT recommended

CRT may be considered

Insufficient evidence for CRT

See Pages 11-12


See next page for continuation (ICD and Patient Preference)


• Routine assessment of dyssynchrony with present echo techniques is not recommended to guide prescription of CRT (see Page 13)

*CRT may be considered for patients in permanent AF who are otherwise suitable for therapy (see Page 14)



Guiding Principles – ICD and Patient Preference






See
Page 18




Guiding Principles – Follow-up and Optimization

Optimize HF therapy throughout patient follow-up



See
Page 21





When to Consider CRT

Recommendation

- CRT is recommended for patients in sinus rhythm with NYHA class II, NYHA class III or ambulatory NYHA class IV heart failure symptoms, a LVEF $\leq 35\%$, and QRS duration ≥ 130 ms because of LBBB
Strong Recommendation, High-quality Evidence

Values and preferences

- This recommendation places great value on the inclusion criteria of the landmark CRT trials, the characteristics of the patients enrolled in these trials, and the derived benefit of CRT in patient groups identified in subsequent analyses of these landmark trials

Practical tip:

- Because these patients were largely excluded in major CRT studies, there is insufficient evidence to recommend CRT in:
 - NYHA class I limitation
 - Nonambulatory class IV NYHA
 - QRS duration < 130 ms
- Patients with LBBB and QRS duration ≥ 150 ms appear more likely to benefit from CRT than patients with non-LBBB conduction and/or QRS prolongation

Recommendations

- CRT may be considered for patients in sinus rhythm with NYHA class II, NYHA class III or ambulatory NYHA class IV heart failure, a LVEF $\leq 35\%$, and QRS duration ≥ 150 ms not because of LBBB conduction

Weak Recommendation, Low-quality Evidence

Practical tip:

- There is no clear evidence of benefit with CRT among patients with QRS duration < 150 ms because of non-LBBB conduction

Values and preferences

- This recommendation places great value on the inclusion criteria of the landmark trials, the characteristics of the patients enrolled, and the quality of this evidence
- This recommendation also places greater weight on QRS duration as a determinant of CRT response

- CRT may be considered for patients with chronic RV pacing or who are likely to be chronically paced, have signs and/or symptoms of heart failure, and a LVEF $\leq 35\%$

Weak Recommendation, Low-quality Evidence

Practical tip:

- Attempts to minimize RV pacing should be undertaken before consideration of CRT upgrade
- There is less evidence for the utility of CRT in patients who do not have pre-existing LBBB and are chronically RV-paced
- Risks of CRT upgrade must be considered and balanced with the potential benefits of CRT upgrade
- Patients undergoing AV junctional ablation with moderate LV dysfunction might benefit from CRT, as may those who have an indication for chronic pacing and characteristics similar to patients randomized in BLOCK HF
- It is often difficult to predict reliably which patients will be chronically RV paced at the time of pacemaker therapy

Values and preferences

- This recommendation places great value on the inclusion criteria of the landmark trials and the characteristics of the patients enrolled

Recommendation

- Routine assessment of dyssynchrony with present echo techniques is not recommended to guide the prescription of CRT

Strong Recommendation, Low-quality Evidence

Values and preferences

- This recommendation takes into account the quality of the evidence and the results of larger, multicentre studies

Practical tip:

- Issues of reproducibility and inter/intra-rater assessment identified in the larger studies limit the routine role of echo to guide the prescription of CRT
- Ongoing imaging research (e.g. scar, viability) is presently under investigation

Recommendation

- CRT may be considered for patients in permanent AF who are otherwise suitable for this therapy
Weak Recommendation, Low-quality Evidence

Values and preferences

- This recommendation places great value on the inclusion criteria of the landmark trials, the characteristics of the patients enrolled, the derived benefit of CRT in patients with permanent AF, and the quality of the evidence

Practical tip:

- The amount of biventricular pacing needs to be evaluated
- Arrhythmia device counters alone might not accurately reflect the true percent of biventricular pacing
- It is important to ensure a very high percentage of biventricular pacing
- AV junctional ablation might be necessary to achieve sufficient biventricular pacing

Recommendation

- A CRT-P is recommended for patients who are suitable for resynchronization therapy, but not for an ICD
Strong Recommendation, Moderate-quality Evidence

Values and preferences

- This recommendation places a greater value on quality of life and improvement of heart failure symptoms, rather than prevention of sudden death

Practical tip:

- CRT-P has been shown to reduce morbidity and mortality in patients with NYHA class III and ambulatory class IV symptoms
- Therapy should be individualized in accordance with the overall goals of care

Recommendation

- In patients treated with CRT, pacing from a nonapical LV epicardial region might be considered
Weak Recommendation, Low-quality Evidence

Values and preferences

- This recommendation places great value on the quality of the evidence



Practical tip:

- Objective evaluation of the pre-CRT implantation functional capacity and symptoms is important, particularly in patients in whom there is disparity between the reported symptoms and the clinical assessment, or to distinguish the non-HF related causes of functional limitation

Recommendation

- We recommend that the prescription of CRT and the choice of platform (CRT-P vs. CRT-D) should take into account clinical factors that would affect the overall goals of care

Strong Recommendation, Moderate-quality Evidence

Values and preferences

- This recommendation places great value on the benefit of CRT in appropriately selected patients with HF, and minimization of risk

Practical tip:

- Comorbid conditions and clinical factors (e.g. age, renal function, frailty) should be considered together, and one alone should not preclude a patient from CRT implantation
- Therapy should be individualized in accordance with the overall goals of care and patient preference

Recommendations

- We suggest that CRT might be considered for patients with new-onset high-degree AV block requiring chronic RV pacing, signs and/or symptoms of HF, and LVEF $\leq 45\%$

Conditional Recommendation, Moderate-quality Evidence

Values and preferences

- This recommendation places value on the knowledge that CRT might provide more benefit than RV apical pacing, even though the strength of evidence is moderate and the available data result in a conditional rating

- We recommend that all patients with HF who are planned to receive a CIED system revision should be considered for their eligibility for upgrade to CRT

Strong Recommendation, Low-Quality Evidence

Values and preferences

- Careful evaluation of risks/benefits of upgrades in CRT-eligible patients with existing CIED systems should be considered

- We suggest that placement of an LV lead at the time of open heart surgery, for the purpose of facilitating future CRT, might be considered in patients for whom CRT is recommended and the need for device therapy is unlikely to be changed by the surgical procedure

Conditional Recommendation, Low-quality Evidence

Values and preferences

- This recommendation places value on practical considerations and multidisciplinary discussion in absence of substantial data

Practical tip:

- Most patients in BLOCK HF had LVEF $\leq 45\%$ and were NYHA class II/III
- BLOCK HF enrolled patients with de novo implants and the same considerations might not apply in patients who are chronically RV-paced
- There is limited RCT data with respect to CRT upgrade and potential benefits must be balanced with the significantly higher risk with CRT vs. generator replacement alone

Practical tip:

- Risk/benefit of adding an LV lead when a procedure is being performed may be favourable

Practical tip:

- Patients with severe structural abnormalities that can be easily corrected with cardiac surgery are the least likely to benefit from LV lead placement
- Balance risk/cost of placing an unnecessary lead with feasibility of placing/testing an LV lead during surgery in an eligible patient
- Avoid apical LV lead position



Recommendation

- We recommend that in patients taking warfarin for whom perioperative anticoagulation is deemed necessary, continued warfarin is recommended over the use of heparin-based bridging
Strong Recommendation, Moderate-quality Evidence

Values and preferences

- This recommendation places great value on safely preventing perioperative bleeding and the quality of the evidence

Practical tip:

- Perioperative OAC is usually deemed necessary in CHADS₂ ≥ 3 or $>5\%$ annual risk
- No data are presently available on the use of novel OAC in this population
 - Interruption of these agents should be of short duration in patients at high risk of thromboembolism, but renal function, procedural risk, and complexity of the procedure might need to be considered in these situations
 - Rapidity of onset and offset make these more facile to control periprocedurally
- Discontinuation of OAC in patients at lower risk of thromboembolism should be considered to minimize the risk of bleeding

Recommendations

- We recommend that CRT implantation be performed only in facilities that have strict infection prevention control standards
Strong Recommendation, Low-quality Evidence

Values and preferences

- This recommendation takes into account that infection prevention is a key aspect to CRT implantation and system revision
- We recommend that appropriate fluoroscopic equipment, radiation shielding, and radiation reduction imaging methods be used to minimize radiation exposure to the operator, patient and other staff
Strong Recommendation, Low-quality Evidence

Values and preferences

- This recommendation takes into account that fluoroscopic exposure poses risk to patients, staff and operators, and must be minimized using all available means

Practical tip:

- In patients with an ischemic etiology or non-LBBB patterns of QRS morphology, correlation of venous anatomy using CMR with mechanical dyssynchrony might be helpful in guiding LV lead placement and optimizing response to CRT
- At the time of implantation the anatomy of the venous system remains the most important limitation (even though concordance between position of LV lead and the area of latest activation seem to predict better response to CRT)
- A secondary analysis of the SMART-AV study revealed Q-LV >95 ms was associated with improved reverse modelling
- Basic determinants of successful LV lead placement remain the same:
 - Pacing with an adequate threshold
 - Long-term stability of the LV lead
 - Avoiding capture of the nearby phrenic nerve
- Emerging technologies to assist in 3-D reconstruction of venous anatomy and its electrical activation patterns are being investigated

Practical tip:

- Define response to CRT using “treatment effect” based on: individual patient goals of improved QoL, symptom reduction, fewer hospitalization and improved survival
 - Include validated composite endpoints (death, hospitalization, QoL)
 - Preferred over more ambiguous “response rate” comprised of a myriad of variables that might not correspond with clinical outcomes or be relevant to CRT recipients

Recommendation

- We recommend that alterations in clinical parameters after vs. before CRT be assessed within 6-12 months after CRT implantation to guide ongoing HF management

Strong Recommendation, Low-quality Evidence

Values and preferences

- This recommendation is based on the importance of ensuring optimal medical therapy and deriving important benefit in improvement in reverse remodelling and QoL with CRT delivery

Practical tip:

- Process of LV remodelling is a dynamic process that begins at time of CRT implantation but requires ongoing reassessment throughout follow-up

Practical tip:

- Suggested starting point: alternating RM visits with direct device clinic follow-up visits in a 1:1 ratio
- Proportion of RM-based follow-up assessments might increase or decrease as dictated by individual patient characteristics



Recommendation

- We suggest that CRT implantation should occur within 6-8 weeks from the decision to implant to avoid preventable adverse events, such as HF hospitalizations and death

Conditional Recommendation, Low-quality Evidence



Canadian Cardiovascular Society

Leadership. Knowledge. Community.



Please visit www.ccs.ca for more information or additional resources.